

Sample Statistical Summary

MAPEP-11-XaW25

Analyte	T(1)	A(2)	Grand(3) Mean	Std Dev	Ref Value	Ref Unc	Acceptance Range
Radiological							Units: (Bq/L)
Iodine-129	27	16	9.4	1.5	9.5	0.1	6.7 - 12.4

Note: (1) T = Total number of laboratories reporting analyte.
 (2) A = Number of laboratories with 'Acceptable' performance.
 (3) Mean excludes values derived as total metals and values indicated as "Not Acceptable".

Other Flags:

A = Result acceptable Bias $\leq 20\%$
 W = Result acceptable with warning $20\% < \text{Bias} < 30\%$
 N = Result not acceptable Bias $> 30\%$
 L = Uncertainty potentially too low (for information purposes only)
 H = Uncertainty potentially too high (for information purposes only)
 RW = Report Warning
 NR = Not Reported